

CERTIFICATE OF ANALYSIS No.: 2023-12955

CLIENT

KANNABIO HEMP HELLAS, SKOUFA 110
38334 VOLOS, Greece

SAMPLE *

COLONAIID



Sample condition: SUITABLE
Sample ID: 2338034
Sample type: Viscous liquid
Batch No.: * K29G

Work order: 2023-107712
Analysis ID: 2023_315
Method ID: PHL_RPC_16C
Method SOP: MET-LAB-001-08

Sample received: 19/09/2023
Start of analysis: 22/09/2023
End of analysis: 26/09/2023
Analyst: Domen Lavriha

* Information provided by the client.

CANNABINOID PROFILE		Concentration [% w/w]	Expanded uncertainty [% w/w]	Graphic presentation of relative cannabinoid concentration
CBDV	- Cannabidivarin	< LOQ	n/a	_____
CBDA	- Cannabidiolic acid	< LOQ	n/a	_____
CBGA	- Cannabigerolic acid	< LOQ	n/a	_____
CBG	- Cannabigerol	10.22	0.72	██████████
CBD	- Cannabidiol	5.92	0.30	██████████
THCV	- Tetrahydrocannavarin	< LOQ	n/a	_____
CBN	- Cannabinol	< LOQ	n/a	_____
Δ⁹-THC	- Δ-9-Tetrahydrocannabinol	< LOQ	n/a	_____
Δ⁸-THC	- Δ-8-Tetrahydrocannabinol	< LOQ	n/a	_____
CBL	- Cannabicyclol	< LOQ	n/a	_____
CBC	- Cannabichromene	2.26	0.11	██████████
Δ⁹-THCA	- Δ-9-Tetrahydrocannabinolic acid	< LOQ	n/a	_____
CBV	- Cannabivarin	< LOQ	n/a	_____
CBCA	- Cannabichromenic acid	< LOQ	n/a	_____
CBT	- Cannabicitran	0.162	0.028	██████████
CBE	- Cannabielsoin	0.037 #	0.010	██████████

Units and abbreviations: % w/w = weight percent, < LOQ = below the limit of quantitation (0.03 % w/w), ND = not detected, n/a = not available.

The results given herein apply only to the sample as received and tested. **Expanded Uncertainty** was calculated using coverage factor $k = 2$, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

26/09/2023

Approved by:



mag. Janja Ahej
Analytical Laboratory Manager

Authorized by:



dr. Boštjan Jančar
Chief Technology Officer

End of Certificate